## § 37.43

Roentgenographic Facility Document for quality control purposes. Approvals granted hereunder may be suspended or withdrawn by notice in writing when in the opinion of ALOSH the quality of roentgenograms or information submitted under this section warrants such action. A copy of a notice withdrawing approval will be sent to each operator who has listed the facility as its facility for giving chest roentgenograms and shall be displayed on the mine bulletin board adjacent to the operator's approved plan. The approved plan will be reevaluated by ALOSH in light of this change.

[43 FR 33715, Aug. 1, 1978; 43 FR 38830, Aug. 31, 1978]

## § 37.43 Protection against radiation emitted by roentgenographic equipment.

Except as otherwise specified in §37.41, roentgenographic equipment, its use and the facilities (including mobile facilities) in which such equipment is used, shall conform to applicable State and Federal regulations (See 21 CFR part 1000). Where no applicable regulations exist, roentgenographic equipment, its use and the facilities (including mobile facilities) in which such equipment is used shall conform to the recommendations of the National Council on Radiation Protection and Measurements in NCRP Report No. 33 "Medical X-ray and Gamma-Ray Protection for Energies up to 10 MeV-Equipment Design and Use" (issued February 1, 1968), in NCRP Report No. 48, "Medical Radiation Protection for Medical and Allied Health Personnel" (issued August 1, 1976), and in NCRP Report No. 49, "Structural Shielding Design and Evaluation for Medical Use of X-rays and Gamma Rays of up to 10 MeV" (issued September 15, 1976). These documents are hereby incorporated by reference and made a part of this subpart. These documents are available for examination at ALOSH, 944 Chestnut Ridge Road, Morgantown, WV 26505, and at the National Institute for Occupational Safety and Health, 5600 Fishers Lane, Rockville, MD 20857. Copies of NCRP Reports Nos. 33, 48, and 49 may be purchased for \$3, \$4.50, and \$3.50 each, respectively, from NCRP

Publications, P.O. Box 30175, Washington, DC 20014.

SPECIFICATIONS FOR INTERPRETATION, CLASSIFICATION, AND SUBMISSION OF CHEST ROENTGENOGRAMS

## § 37.50 Interpreting and classifying chest roentgenograms.

- (a) Chest roentgenograms shall be interpreted and classified in accordance with the ILO Classification system and recorded on a Roentgenographic Interpretation Form (Form CDC/NIOSH (M)2.8).
- (b) Roentgenograms shall be interpreted and classified only by a physician who regularly reads chest roentgenograms and who has demonstrated proficiency in classifying the pneumoconioses in accordance with §37.51.
- (c) All interpreters, whenever interpreting chest roentgenograms made under the Act, shall have immediately available for reference a complete set of the ILO International Classification of Radiographs for Pneumoconioses, 1980.

NOTE: This set is available from the International Labor Office, 1750 New York Avenue, NW., Washington, DC 20006 (Phone: 202/376-2315).

- (d) In all view boxes used for making interpretations:
- (1) Fluorescent lamps shall be simultaneously replaced with new lamps at 6-month intervals;
- (2) All the fluorescent lamps in a panel of boxes shall have identical manufacturer's ratings as to intensity and color:
- (3) The glass, internal reflective surfaces, and the lamps shall be kept clean:
- (4) The unit shall be so situated as to minimize front surface glare.

[43 FR 33715, Aug. 1, 1978, as amended at 49 FR 7564, Mar. 1, 1984]

## § 37.51 Proficiency in the use of systems for classifying the pneumoconioses.

- (a) First or "A" readers:
- (1) Approval as an "A" reader shall continue if established prior to (insert) effective date of these regulations).